

TMIC Li-Node Service Sample Amount Requirement

TMIC Li-Node provides Global Metabolomics Analysis by using Chemical Isotope Labeling (CIL) LC-MS platform and Global Lipidomics Analysis by using In-depth Lipidomics LC-MS platform. The sample amount required for each level of analysis is listed below. If the sample type is not listed or the sample amount is not sufficient as shown below, please contact TMIC Li-Node (liang.li@ualberta.ca). We will discuss and try to accommodate your needs. In some cases, we may need a few test samples for method development.

Sample Type	Basic metabolomics ¹	Elevated metabolomics ¹	Comprehensive metabolomics ¹	Flexible metabolomics	In-depth lipidomics ^{1,4}	Both metabolomics and lipidomics
Serum/Plasma	50 µL / 100 µL	90 µL / 150 µL	180 µL / 250 µL	60 µL/channel	10 µL / 20 µL	In one tube
Urine	15 µL / 30 µL	25 µL / 50 µL	40 µL / 80 µL	20 µL/channel	240 µL / 500 µL	In one tube
Saliva	50 µL / 100 µL	90 µL / 150 µL	180 µL / 250 µL	60 µL/channel	60 µL / 120 µL	In one tube
Other Common Biofluids ²	50 µL / 100 µL	90 µL / 150 µL	180 µL / 250 µL	60 µL/channel	60 µL / 120 µL	In one tube
Tissue ²	10 mg / 20 mg	20 mg / 40 mg	40 mg / 80 mg	20 mg/channel	2 mg / 5 mg	Preferably placed in different tubes
Feces	10 mg / 20 mg	20 mg / 40 mg	40 mg / 80 mg	20 mg/channel	2 mg / 5 mg	Preferably placed in different tubes
Plant	10 mg / 20 mg	20 mg / 40 mg	40 mg / 80 mg	20 mg/channel	2 mg / 5 mg	Preferably placed in different tubes
Cells	0.5 × 10 ⁶ / 1 × 10 ⁶	1 × 10 ⁶ / 2 × 10 ⁶	2 × 10 ⁶ / 4 × 10 ⁶	1 × 10 ⁶ /channel	2 × 10 ⁶ / 4 × 10 ⁶	Preferably placed in different tubes
Microbes	20 mg / 40 mg	20 mg / 60 mg	20 mg / 80 mg	40 mg/channel	2 mg / 5 mg	Preferably placed in different tubes
Culture Medium	50 µL / 100 µL	90 µL / 150 µL	180 µL / 250 µL	100 µL/channel	240 µL / 500 µL	In one tube
Other Types	Please inquire by sending an email to liang.li@ualberta.ca					

1. The sample amount in each cell is shown as “**minimum amount / preferred amount**”.
2. Not including special biofluids or tissues, such as much diluted samples (e.g., dialysates, exosomes, needle biopsy, etc.). Please contact us for discussion.
3. Please measure or weigh the amount of samples during sample collection. Send the sample amount information when sending the samples.
4. For lipidomics of solid samples (e.g., tissue, cell pellets, etc.), purging the samples with N₂ or Ar before sealing the vials for shipping is recommended.